

Office of the Vice President and
Chief Financial Officer
Finance and Accounting Division
Purchasing and Disbursement Services
<http://purchasing.ufl.edu/>

102 Elmore Hall
PO Box 115250
Gainesville, FL 32611-5250
352-392-1331
352-392-8837 Fax

July 16, 2009

ADDENDUM NUMBER 4 ON INVITATION TO BID# ITB09MN-122

The following changes to the Substrates Specifications have been made to ITB09MN-122:

1. For PRM, F-PRM, SRM and F-SRM mirror substrates:

Error, High Spatial Frequency: 1– 750 mm⁻¹ (page 3 of each specification)

Surfaces 1 and 2 - HSF error $\sigma_{\text{rms}} \leq 0.2$ nanometers

Measured at the following locations:

1. Within 2mm of the center of the surface.
2. Four positions equally spaced along the circumference of a centered, 10 mm diameter circle.
3. Three positions equally spaced along the circumference of a centered, 20 mm diameter circle.

2. For PR2, F-PR2 and F-SR2 mirror substrates:

Error, High Spatial Frequency: 1– 750 mm⁻¹ (page 3 of each specification)

Surface 1- HSF error $\sigma_{\text{rms}} \leq 0.2$ nanometers

Measured at the following locations:

1. Within 2mm of the center of the surface.
2. Four positions equally spaced along the circumference of a centered, 10 mm diameter circle.
3. Three positions equally spaced along the circumference of a centered, 20 mm diameter circle.

3. For SR2 mirror substrates:

Error, High Spatial Frequency: 1– 750 mm⁻¹ (page 3 of each specification)

Surface 1- HSF error $\sigma_{\text{rms}} \leq 0.2$ nanometers

Measured at the following locations:

1. Within 2mm of the center of the surface.
2. Four positions equally spaced along the circumference of a centered, 10 mm diameter circle.
3. **Five** positions equally spaced along the circumference of a centered, 20 mm diameter circle.

The revised Substrate Specifications and Drawings package including these changes will follow in a separate addendum. The new Specifications and Drawings versions will be -v3-.

This addendum shall be considered part of the Contract Documents for the above mentioned project as though it had been issued at the same time and incorporated integrally therewith. Where provisions of the following supplementary data differ from those of the original Contract documents, this addendum shall govern and take precedence. Bidders are hereby notified that they shall acknowledge receipt of the addendum.

Sincerely,

A handwritten signature in blue ink, appearing to read "Michael Neff".

Michael Neff
Purchasing Coordinator, II

PLEASE ACKNOWLEDGE RECEIPT OF ADDENDUM AND RETURN WITH YOUR BID. FAILURE TO ACKNOWLEDGE YOUR ADDENDUM COULD CONSTITUTE REJECTION OF YOUR BID.

VENDOR NAME

VENDOR ADDRESS

SIGNATURE