

Office of the Vice President
and Chief Financial Officer
Finance and Accounting Division
Purchasing and Disbursement Services
<http://purchasing.ufl.edu/>

102 Elmore Hall
PO Box 115250
Gainesville, FL 32611-5250
352-392-1331
352-392-8837 Fax

August 4, 2009

ADDENDUM NUMBER 1 ON INVITATION TO BID# ITB10MN-109

Clarification of Specifications:

The microscope requires a secondary electron detector. The type of detector required is a conventional secondary electron detector called an E-T (Everhart-Thornley) in chamber detector.

Backscatter Electron Detector and Energy Dispersive Spectrometer are not part of the specifications for this bid.

This addendum shall be considered part of the Contract Documents for the above mentioned project as though it had been issued at the same time and incorporated integrally therewith. Where provisions of the following supplementary data differ from those of the original Contract documents, this addendum shall govern and take precedence. Bidders are hereby notified that they shall acknowledge receipt of the addendum.

Sincerely,



Michael Neff
Purchasing Coordinator, II

PLEASE ACKNOWLEDGE RECEIPT OF ADDENDUM AND RETURN WITH YOUR BID. FAILURE TO ACKNOWLEDGE YOUR ADDENDUM COULD CONSTITUTE REJECTION OF YOUR BID.

VENDOR NAME

VENDOR ADDRESS

SIGNATURE